

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051319\
 Data File : VX009534.D
 Acq On : 13 May 2019 11:15
 Operator : JC/SP
 Sample : VX0513WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0513WBL01

Quant Time: May 14 03:56:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	180066	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	284374	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	247803	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	104015	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	129713	52.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.54%	
35) Dibromofluoromethane	5.49	113	96574	54.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.04%	
50) Toluene-d8	8.71	98	390643	54.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.26%	
62) 4-Bromofluorobenzene	11.13	95	126810	48.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.70%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051319\
Data File : VX009534.D
Acq On : 13 May 2019 11:15
Operator : JC/SP
Sample : VX0513WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0513WBL01

Quant Time: May 14 03:56:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX009534.D

Acq: 13 May 2019 11:15

Instrument :

MSVOA_X

ClientSampleId :


VX0513WBL01

Tot Ion:168 Resp: 180066

Ion Ratio Lower Upper




168 100

99 54.3 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 52.769 ug/l

RT: 6.06 min Scan# 804

Delta R.T. -0.01 min

Lab File: VX009534.D


Acq: 13 May 2019 11:15

Tgt Ion: 65 Resp: 129713

Ion Ratio Lower Upper



65 100


67 53.2 0.0 106.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.85 min Scan# 935
 Delta R.T. -0.01 min
 Lab File: VX009534.D
 Acq: 13 May 2019 11:15

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0513WBL01

Tot Ion:114 Resp: 284374
 Ion Ratio Lower Upper
 114 100
 63 23.6 0.0 45.6
 88 17.1 0.0 33.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 54.022 ug/l
 RT: 5.49 min Scan# 712
 Delta R.T. 0.00 min
 Lab File: VX009534.D
 Acq: 13 May 2019 11:15

Tgt Ion:113 Resp: 96574
 Ion Ratio Lower Upper
 113 100
 111 103.1 82.8 124.2
 192 21.9 17.0 25.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 54.126 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX009534.D

Acq: 13 May 2019 11:15

Instrument :

MSVOA_X

ClientSampleId :


VX0513WBL01

Tot Ion: 98 Resp: 390643

Ion Ratio Lower Upper




98 100

100 64.4 51.2 76.8



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 48.354 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX009534.D

Acq: 13 May 2019 11:15


Tgt Ion: 95 Resp: 126810

Ion Ratio Lower Upper

95 100

174 83.8 0.0 161.6



176 80.2 0.0 159.2




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX009534.D
Acq: 13 May 2019 11:15

Instrument :
MSVOA_X
ClientSampleId :
VX0513WBL01

Tot Ion:117 Resp: 247803
Ion Ratio Lower Upper
117 100
82 59.3 48.8 73.2
119 32.1 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX009534.D
Acq: 13 May 2019 11:15

Tgt Ion:152 Resp: 104015
Ion Ratio Lower Upper
152 100
115 59.6 41.2 123.6
150 158.5 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

