

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051319\
 Data File : VX009536.D
 Acq On : 13 May 2019 12:25
 Operator : JC/SP
 Sample : K2834-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-TB-20190509

Quant Time: May 14 04:00:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	184556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	291709	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	256460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112922	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118732	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
35) Dibromofluoromethane	5.50	113	88773	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
50) Toluene-d8	8.71	98	364455	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.13	95	122128	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
Target Compounds						
16) Acetone	2.44	43	9584	7.084	ug/l	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051319\
Data File : VX009536.D
Acq On : 13 May 2019 12:25
Operator : JC/SP
Sample : K2834-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB202-TB-20190509

Quant Time: May 14 04:00:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX009536.D

Acq: 13 May 2019 12:25

Instrument :

MSVOA_X

ClientSampleId :


BP-VPB202-TB-20190509

Tot Ion:168 Resp: 184556

Ion Ratio Lower Upper

168 100

99 53.3 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80 5.90



#16

Acetone

Concen: 7.084 ug/l

RT: 2.44 min Scan# 211

Delta R.T. -0.01 min

Lab File: VX009536.D


Acq: 13 May 2019 12:25

Tgt Ion: 43 Resp: 9584

Ion Ratio Lower Upper

43 100

58 30.5 24.9 37.3



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.44

6000

5000

4000


3000

2000

1000

0

Time--> 2.30 2.40 2.50









#33
 1,2-Dichloroethane-d4
 Concen: 47.127 ug/l
 RT: 6.06 min Scan# 805
 Delta R.T. 0.00 min
 Lab File: VX009536.D
 Acq: 13 May 2019 12:25




Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-TB-20190509


Tot Ion: 65 Resp: 118732
 Ion Ratio Lower Upper
 65 100
 67 53.8 0.0 106.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX009536.D
 Acq: 13 May 2019 12:25

Tgt Ion: 114 Resp: 291709
 Ion Ratio Lower Upper
 114 100
 63 22.8 0.0 45.6
 88 16.5 0.0 33.2









#35
 Dibromofluoromethane
 Concen: 48.409 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.01 min
 Lab File: VX009536.D
 Acq: 13 May 2019 12:25




Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-TB-20190509


Tot Ion:113 Resp: 88773
 Ion Ratio Lower Upper
 113 100
 111 102.0 82.8 124.2
 192 22.2 17.0 25.6



#50
 Toluene-d8
 Concen: 49.227 ug/l
 RT: 8.71 min Scan# 1240
 Delta R.T. -0.00 min
 Lab File: VX009536.D
 Acq: 13 May 2019 12:25

Tgt Ion: 98 Resp: 364455
 Ion Ratio Lower Upper
 98 100
 100 64.5 51.2 76.8





#62

4-Bromofluorobenzene

Concen: 45.398 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX009536.D

Acq: 13 May 2019 12:25

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB202-TB-20190509


Tot Ion: 95 Resp: 122128

Ion Ratio Lower Upper

95 100

174 82.7 0.0 161.6




176 80.5 0.0 159.2



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1469

Delta R.T. 0.00 min

Lab File: VX009536.D

Acq: 13 May 2019 12:25


Tgt Ion: 117 Resp: 256460

Ion Ratio Lower Upper

117 100

82 59.6 48.8 73.2



119 32.0 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX009536.D

Acq: 13 May 2019 12:25

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB202-TB-20190509


Tot Ion:152 Resp: 112922

Ion Ratio Lower Upper

152 100

115 58.2 41.2 123.6

150 155.5 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

