

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051320\
 Data File : VX016169.D
 Acq On : 13 May 2020 11:19
 Operator : JC/SP
 Sample : L2597-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-TB4-200511

Quant Time: May 14 06:37:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	295847	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	478104	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	465771	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	231496	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	175476	45.07	ug/l	0.00
Spiked Amount						
			Recovery	=	90.14%	
35) Dibromofluoromethane	5.47	113	140423	46.36	ug/l	0.00
Spiked Amount						
			Recovery	=	92.72%	
50) Toluene-d8	8.70	98	596135	51.08	ug/l	0.00
Spiked Amount						
			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	11.12	95	234469	52.82	ug/l	0.00
Spiked Amount						
			Recovery	=	105.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051320\
Data File : VX016169.D
Acq On : 13 May 2020 11:19
Operator : JC/SP
Sample : L2597-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-TB4-200511

Quant Time: May 14 06:37:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

