

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051320\
 Data File : VX016175.D
 Acq On : 13 May 2020 13:35
 Operator : JC/SP
 Sample : L2597-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW592-200511

Quant Time: May 14 06:50:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	281249	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	458435	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	459840	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	233463	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	170640	46.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.22%	
35) Dibromofluoromethane	5.47	113	137940	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
50) Toluene-d8	8.70	98	574774	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
62) 4-Bromofluorobenzene	11.12	95	230965	54.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.52%	
Target Compounds						
16) Acetone	2.42	43	3137	1.900	ug/l	97
44) Trichloroethene	7.19	130	4503	0.960	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051320\
Data File : VX016175.D
Acq On : 13 May 2020 13:35
Operator : JC/SP
Sample : L2597-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW592-200511

Quant Time: May 14 06:50:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

