

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051320\
 Data File : VX016180.D
 Acq On : 13 May 2020 15:28
 Operator : JC/SP
 Sample : L2597-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW720-200511

Quant Time: May 14 07:00:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	279438	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	466718	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	457254	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	237869	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	169366	46.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.12%	
35) Dibromofluoromethane	5.47	113	136410	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
50) Toluene-d8	8.70	98	574919	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
62) 4-Bromofluorobenzene	11.12	95	230988	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
Target Compounds						
16) Acetone	2.42	43	3074	1.874	ug/l	94
27) cis-1,2-Dichloroethene	4.57	96	4251	1.201	ug/l	79
44) Trichloroethene	7.19	130	1291	0.270	ug/l	89
64) Tetrachloroethene	9.32	164	1303	0.232	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051320\
Data File : VX016180.D
Acq On : 13 May 2020 15:28
Operator : JC/SP
Sample : L2597-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW720-200511

Quant Time: May 14 07:00:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

