

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051320\
 Data File : VX016181.D
 Acq On : 13 May 2020 15:51
 Operator : JC/SP
 Sample : L2597-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW721-200511

Quant Time: May 14 07:02:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	279060	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	462358	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	457117	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	233602	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	170698	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
35) Dibromofluoromethane	5.47	113	135765	46.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.70%	
50) Toluene-d8	8.70	98	570257	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
62) 4-Bromofluorobenzene	11.12	95	227253	52.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.88%	
Target Compounds						
3) Chloromethane	1.32	50	1613	0.470	ug/l	97
16) Acetone	2.42	43	4493	2.742	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051320\
Data File : VX016181.D
Acq On : 13 May 2020 15:51
Operator : JC/SP
Sample : L2597-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW721-200511

Quant Time: May 14 07:02:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

