

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051320\
 Data File : VX016186.D
 Acq On : 13 May 2020 17:44
 Operator : JC/SP
 Sample : L2626-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW732R-200512

Quant Time: May 14 07:10:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	270714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	447433	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	440184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	226494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	168729	47.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.72%	
35) Dibromofluoromethane	5.47	113	134145	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
50) Toluene-d8	8.70	98	549650	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	11.12	95	219635	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	2030	0.721	ug/l #	75
16) Acetone	2.42	43	5138	3.233	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	2502	0.730	ug/l	78
40) Benzene	6.12	78	4124	0.320	ug/l	97
64) Tetrachloroethene	9.32	164	2091	0.387	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051320\
Data File : VX016186.D
Acq On : 13 May 2020 17:44
Operator : JC/SP
Sample : L2626-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1