

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051320\
 Data File : VX016188.D
 Acq On : 13 May 2020 18:29
 Operator : JC/SP
 Sample : L2626-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW730-200512

Quant Time: May 14 07:14:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	264487	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	444565	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	438524	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228922	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	164389	47.23	ug/l	0.00
Spiked Amount			Recovery	=	94.46%	
35) Dibromofluoromethane	5.47	113	130322	46.27	ug/l	0.00
Spiked Amount			Recovery	=	92.54%	
50) Toluene-d8	8.70	98	546280	50.34	ug/l	0.00
Spiked Amount			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.12	95	221546	53.67	ug/l	0.00
Spiked Amount			Recovery	=	107.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.39	62	6733	1.941	ug/l	99
12) 1,1-Dichloroethene	2.36	96	140133	50.969	ug/l	98
16) Acetone	2.42	43	2108	1.357	ug/l	94
19) Methyl tert-butyl Ether	3.17	73	5101	0.538	ug/l	95
21) trans-1,2-Dichloroethene	3.15	96	2420	0.790	ug/l #	76
24) 1,1-Dichloroethane	3.67	63	9713	1.826	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	80477	24.027	ug/l	89
42) 1,2-Dichloroethane	6.17	62	3749	0.813	ug/l	98
44) Trichloroethene	7.19	130	25301	5.562	ug/l	93
49) 1,4-Dioxane	7.72	88	2613	28.412	ug/l	95
64) Tetrachloroethene	9.32	164	166142	30.862	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051320\
Data File : VX016188.D
Acq On : 13 May 2020 18:29
Operator : JC/SP
Sample : L2626-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW730-200512

Quant Time: May 14 07:14:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

