

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051320\
 Data File : VX016216.D
 Acq On : 14 May 2020 06:10
 Operator : JC/SP
 Sample : L2626-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-DUP-0555-200512

Quant Time: May 14 08:04:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	270860	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	445358	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	439246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228221	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	165537	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
35) Dibromofluoromethane	5.46	113	132661	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
50) Toluene-d8	8.70	98	556532	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	11.12	95	221032	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.32	50	1585	0.476	ug/l	89
12) 1,1-Dichloroethene	2.36	96	5614	1.994	ug/l	94
16) Acetone	2.42	43	4537	2.853	ug/l	98
44) Trichloroethene	7.19	130	1299	0.285	ug/l	88
64) Tetrachloroethene	9.32	164	6644	1.232	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051320\
Data File : VX016216.D
Acq On : 14 May 2020 06:10
Operator : JC/SP
Sample : L2626-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-DUP-0555-200512

Quant Time: May 14 08:04:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

