


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX051320\
 Data File : VX016165.D
 Acq On : 13 May 2020 09:21
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Title : SW846 8260
 Last Update : Wed May 06 06:43:32 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	11101	PASS
75	95	30	60	48.1	31336	PASS
95	95	100	100	100.0	65141	PASS
96	95	5	9	6.4	4183	PASS
173	174	0.00	2	1.0	539	PASS
174	95	50	100	82.8	53946	PASS
175	174	5	9	7.5	4048	PASS
176	174	95	101	96.7	52189	PASS
177	176	5	9	6.6	3468	PASS