

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021938.D
 Acq On : 12 May 2021 19:43
 Operator : JC/MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX051321

Quant Time: May 13 06:16:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 13 06:11:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	19482	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	109955	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	97925	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	44273	29.50	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	98.33%		
60) 4-Bromofluorobenzene	11.085	95	48110	29.60	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	98.67%		
63) Toluene-d8	8.653	98	133164	29.76	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	99.20%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	21206	18.44	ug/l	93
3) Chloromethane	1.294	50	24534	17.76	ug/l	99
4) Vinyl Chloride	1.374	62	26038	18.15	ug/l	97
5) Bromomethane	1.611	94	14283	17.69	ug/l	99
6) Chloroethane	1.691	64	16518	17.88	ug/l	95
7) Trichlorofluoromethane	1.892	101	34926	17.42	ug/l	98
8) Diethyl Ether	2.136	74	14932	17.37	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.331	101	20047	17.44	ug/l	95
10) 1,1-Dichloroethene	2.325	96	19720	17.03	ug/l	95
11) Methyl Iodide	2.453	142	22533	19.57	ug/l	95
12) Methyl Acetate	2.709	43	29005	18.11	ug/l	97
13) Acrolein	2.239	56	24132	97.18	ug/l	95
14) Acrylonitrile	3.069	53	57969	88.73	ug/l	97
15) Acetone	2.386	58	17228	88.33	ug/l	90
16) Carbon Disulfide	2.514	76	56792	17.93	ug/l	97
17) Allyl chloride	2.666	41	35594	17.44	ug/l	99
18) Methylene Chloride	2.794	84	24051	17.47	ug/l	94
19) trans-1,2-Dichloroethene	3.099	96	22337	17.75	ug/l	98
20) Diisopropyl ether	3.770	45	66513	17.29	ug/l	93
21) 1,1-Dichloroethane	3.617	63	41094	17.83	ug/l	98
22) cis-1,2-Dichloroethene	4.495	96	26329	17.80	ug/l	97
23) tert-Butyl Alcohol	2.977	59	27053	85.18	ug/l #	100
24) Methyl tert-Butyl Ether	3.117	73	67674	17.95	ug/l	99
25) Chloroform	5.105	83	41529	17.78	ug/l	98
26) Cyclohexane	5.483	56	33232	17.48	ug/l #	97
29) 1,1-Dichloropropene	5.702	75	31113	18.08	ug/l	98
30) 2-Butanone	4.562	43	87860	90.19	ug/l	100
31) 2,2-Dichloropropane	4.489	77	35191	17.40	ug/l	98
32) 1,1,1-Trichloroethane	5.397	97	36129	17.79	ug/l	98
33) Carbon Tetrachloride	5.684	117	31453	18.25	ug/l	91
34) Benzene	6.044	78	94578	18.01	ug/l	98
35) Methacrylonitrile	4.928	41	18680	17.91	ug/l	95
36) 1,2-Dichloroethane	6.092	62	32638	18.21	ug/l #	88
37) Trichloroethene	7.135	130	23279	17.64	ug/l	99
38) Methylcyclohexane	7.391	83	34784	17.80	ug/l	99
39) 1,2-Dichloropropane	7.434	63	23173	17.68	ug/l	92
40) Dibromomethane	7.586	93	16574	18.17	ug/l	95
41) Bromodichloromethane	7.830	83	32974	18.10	ug/l	97
42) Vinyl Acetate	3.727	43	328483	90.30	ug/l	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021938.D
 Acq On : 12 May 2021 19:43
 Operator : JC/MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX051321

Quant Time: May 13 06:16:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 13 06:11:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	4.721	43	37059	18.72	ug/l	100
44) Isopropyl Acetate	6.348	43	59127	18.11	ug/l	100
45) 1,4-Dioxane	7.677	88	10181	349.28	ug/l #	100
46) Methyl methacrylate	7.696	41	27914	18.44	ug/l	93
47) n-amyl Acetate	10.848	43	48414	17.83	ug/l	99
48) t-1,3-Dichloropropene	8.982	75	36859	18.03	ug/l	94
49) cis-1,3-Dichloropropene	8.372	75	40139	18.29	ug/l	98
50) 1,1,2-Trichloroethane	9.159	97	23420	18.04	ug/l	96
51) Ethyl methacrylate	9.122	69	37440	17.88	ug/l	97
52) 1,3-Dichloropropane	9.311	76	40602	17.94	ug/l	98
53) Dibromochloromethane	9.525	129	24186	18.54	ug/l	97
54) 1,2-Dibromoethane	9.616	107	24696	18.52	ug/l	99
55) 2-Chloroethyl vinyl ether	8.244	63	103898	94.00	ug/l	99
56) Bromoform	10.805	173	14903	18.01	ug/l	97
58) 4-Methyl-2-Pentanone	8.580	43	178103	90.19	ug/l	99
59) 2-Hexanone	9.433	43	132682	90.72	ug/l	98
61) Tetrachloroethene	9.275	164	19690	17.73	ug/l	85
62) Toluene	8.726	91	99848	17.81	ug/l	99
64) Chlorobenzene	10.086	112	61809	18.44	ug/l	99
65) 1,1,1,2-Tetrachloroethane	10.165	131	21708	17.74	ug/l	97
66) Ethyl Benzene	10.195	91	111914	18.14	ug/l	100
67) m/p-Xylenes	10.305	106	83292	35.63	ug/l	98
68) o-Xylene	10.646	106	41688	18.31	ug/l	99
69) Styrene	10.659	104	67543	17.73	ug/l	99
70) Isopropylbenzene	10.963	105	105799	17.69	ug/l	98
71) 1,1,2,2-Tetrachloroethane	11.213	83	34784	17.64	ug/l	98
72) 1,2,3-Trichloropropane	11.244	75	41284	23.62	ug/l	63
73) Bromobenzene	11.201	156	21907	17.29	ug/l	94
74) n-propylbenzene	11.305	91	123364	17.70	ug/l	100
75) 2-Chlorotoluene	11.366	91	74083	17.94	ug/l	99
76) 1,3,5-Trimethylbenzene	11.457	105	89390	17.90	ug/l	99
77) t-1,4-Dichloro-2-butene	11.024	75	12800	18.07	ug/l	93
78) 4-Chlorotoluene	11.457	91	83674	17.61	ug/l	98
79) tert-butylbenzene	11.719	119	86143	18.04	ug/l	99
80) 1,2,4-Trimethylbenzene	11.756	105	88713	17.78	ug/l	99
81) sec-Butylbenzene	11.896	105	105444	17.57	ug/l	100
82) p-Isopropyltoluene	12.012	119	88105	17.51	ug/l	98
83) 1,3-Dichlorobenzene	11.975	146	41566	17.31	ug/l	98
84) 1,4-Dichlorobenzene	12.043	146	42443	17.62	ug/l	98
85) n-Butylbenzene	12.335	91	78695	17.06	ug/l	99
86) Hexachloroethane	12.542	117	15161	18.13	ug/l	98
87) 1,2-Dichlorobenzene	12.341	146	41132	17.68	ug/l	100
88) 1,2-Dibromo-3-Chloropr...	12.945	75	7529	17.15	ug/l	93
89) 1,2,4-Trichlorobenzene	13.591	180	24109	16.92	ug/l	96
90) Hexachlorobutadiene	13.731	225	9892	17.94	ug/l	94
91) Naphthalene	13.780	128	98807	17.68	ug/l	99
92) 1,2,3-Trichlorobenzene	13.963	180	25010	17.68	ug/l	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 30.00 ug/l
RT: 4.910 min Scan# 627
Delta R.T. 0.001 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:128 Resp: 19482
Ion Ratio Lower Upper
128 100
49 168.5 0.0 442.5
130 125.1 0.0 323.3



#2
Dichlorodifluoromethane
Concen: 18.44 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 85 Resp: 21206
Ion Ratio Lower Upper
85 100
87 33.5 14.9 44.5









#3
Chloromethane
Concen: 17.76 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 50 Resp: 24534
Ion Ratio Lower Upper
50 100
52 32.0 26.2 39.4



#4
Vinyl Chloride
Concen: 18.15 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 62 Resp: 26038
Ion Ratio Lower Upper
62 100
64 32.5 24.9 37.3









#5
Bromomethane
Concen: 17.69 ug/l
RT: 1.611 min Scan# 86
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 94 Resp: 14283
Ion Ratio Lower Upper
94 100
96 96.1 76.1 114.1



#6
Chloroethane
Concen: 17.88 ug/l
RT: 1.691 min Scan# 99
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 64 Resp: 16518
Ion Ratio Lower Upper
64 100
66 28.3 24.9 37.3









#7
Trichlorofluoromethane
Concen: 17.42 ug/l
RT: 1.892 min Scan# 132
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:101 Resp: 34926
Ion Ratio Lower Upper
101 100
103 63.9 52.4 78.6



#8
Diethyl Ether
Concen: 17.37 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 74 Resp: 14932
Ion Ratio Lower Upper
74 100
45 101.5 21.1 189.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 17.44 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:101 Resp: 20047
Ion Ratio Lower Upper
101 100
85 43.4 0.0 93.0
151 64.6 0.0 136.6



#10
1,1-Dichloroethene
Concen: 17.03 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 96 Resp: 19720
Ion Ratio Lower Upper
96 100
61 160.1 121.0 181.4
98 62.3 49.1 73.7





#11

Methyl Iodide

Concen: 19.57 ug/l

RT: 2.453 min Scan# 224

Delta R.T. -0.006 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43





Tgt Ion:142 Resp: 22533

Ion Ratio Lower Upper

142 100

127 39.5 21.3 63.7

141 12.1 7.2 21.6



#12

Methyl Acetate

Concen: 18.11 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.000 min

Lab File: VX021938.D




Acq: 12 May 2021 19:43


Tgt Ion: 43 Resp: 29005

Ion Ratio Lower Upper

43 100

74 24.6 18.5 27.7









#13
Acrolein
Concen: 97.18 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 56 Resp: 24132
Ion Ratio Lower Upper
56 100
55 68.1 57.6 86.4



#14
Acrylonitrile
Concen: 88.73 ug/l
RT: 3.069 min Scan# 325
Delta R.T. 0.001 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 53 Resp: 57969
Ion Ratio Lower Upper
53 100
52 82.2 39.3 117.9
51 35.1 17.1 51.2









#15
Acetone
Concen: 88.33 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.001 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 58 Resp: 17228
Ion Ratio Lower Upper
58 100
43 290.9 249.5 374.3



#16
Carbon Disulfide
Concen: 17.93 ug/l
RT: 2.514 min Scan# 234
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 76 Resp: 56792
Ion Ratio Lower Upper
76 100
78 9.4 6.7 10.1





#17

Allyl chloride

Concen: 17.44 ug/l

RT: 2.666 min Scan# 259





Delta R.T. -0.006 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43

Tgt Ion:	41	Resp:	35594
Ion Ratio	Lower	Upper	
41	100		
39	67.6	34.5	103.5
76	39.3	19.8	59.3

Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321



#18

Methylene Chloride

Concen: 17.47 ug/l




RT: 2.794 min Scan# 280


Delta R.T. 0.000 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43

Tgt Ion:	84	Resp:	24051
Ion Ratio	Lower	Upper	
84	100		
49	115.0	101.6	152.4
51	37.7	30.5	45.7
86	65.8	53.4	80.2









#19
trans-1,2-Dichloroethene
Concen: 17.75 ug/l
RT: 3.099 min Scan# 330
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 96 Resp: 22337
Ion Ratio Lower Upper
96 100
61 138.4 113.0 169.4
98 63.3 51.7 77.5



#20
Diisopropyl ether
Concen: 17.29 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 45 Resp: 66513
Ion Ratio Lower Upper
45 100
43 66.5 58.5 87.7
87 31.1 22.2 33.4
59 13.0 9.0 13.6









#21
1,1-Dichloroethane
Concen: 17.83 ug/l
RT: 3.617 min Scan# 415
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:	63	Resp:	41094
Ion Ratio	Lower	Upper	
63	100		
98	8.3	3.8	11.4
100	6.1	2.9	8.6



#22
cis-1,2-Dichloroethene
Concen: 17.80 ug/l
RT: 4.495 min Scan# 559
Delta R.T. -0.006 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:	96	Resp:	26329
Ion Ratio	Lower	Upper	
96	100		
61	139.1	114.0	171.0
98	64.1	53.0	79.6









#23
tert-Butyl Alcohol
Concen: 85.18 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.006 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 59 Resp: 27053
Ion Ratio Lower Upper
59 100
52 0.0 0.0 0.0



#24
Methyl tert-Butyl Ether
Concen: 17.95 ug/l
RT: 3.117 min Scan# 333
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 73 Resp: 67674
Ion Ratio Lower Upper
73 100
43 19.7 16.3 24.5





#25

Chloroform

Concen: 17.78 ug/l

RT: 5.105 min Scan# 659

Delta R.T. 0.000 min

Lab File: VX021938.D





Acq: 12 May 2021 19:43

Tgt Ion: 83 Resp: 41529

Ion Ratio Lower Upper

83 100

85 63.1 51.9 77.9



#26

Cyclohexane

Concen: 17.48 ug/l

RT: 5.483 min Scan# 721

Delta R.T. 0.006 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43




Tgt Ion: 56 Resp: 33232


Ion Ratio Lower Upper

56 100

89 0.0 0.4 0.6#

84 92.4 71.8 107.6





#27

1,2-Dichloroethane-d4

Concen: 29.50 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43

Tgt Ion: 65 Resp: 44273

Ion Ratio Lower Upper

65 100





67 53.1 41.5 62.3

Instrument :

MSVOA_X

ClientSampleId :

ICVVX051321



#28

1,4-Difluorobenzene

Concen: 30.00 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43




Tgt Ion: 114 Resp: 109955


Ion Ratio Lower Upper

114 100

63 20.2 16.6 25.0

88 17.0 13.1 19.7





#29

1,1-Dichloropropene

Concen: 18.08 ug/l

RT: 5.702 min Scan# 757

Delta R.T. 0.000 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43

Tgt Ion: 75 Resp: 31113

Ion Ratio Lower Upper

75 100

110 35.1 0.0 71.4





77 29.5 0.0 63.6

Instrument :

MSVOA_X

ClientSampleId :

ICVVX051321



#30

2-Butanone

Concen: 90.19 ug/l

RT: 4.562 min Scan# 570

Delta R.T. 0.000 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43




Tgt Ion: 43 Resp: 87860


Ion Ratio Lower Upper

43 100

57 7.8 6.4 9.6

72 27.5 22.1 33.1









#31
2,2-Dichloropropane
Concen: 17.40 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 77 Resp: 35191
Ion Ratio Lower Upper
77 100
97 21.9 18.1 27.1



#32
1,1,1-Trichloroethane
Concen: 17.79 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 97 Resp: 36129
Ion Ratio Lower Upper
97 100
99 65.9 51.9 77.9
61 44.0 36.3 54.5









#33
Carbon Tetrachloride
Concen: 18.25 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:117 Resp: 31453
Ion Ratio Lower Upper
117 100
119 89.3 80.8 121.2
121 30.0 23.8 35.8



#34
Benzene
Concen: 18.01 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 78 Resp: 94578
Ion Ratio Lower Upper
78 100
77 22.9 17.8 26.8
51 14.7 10.6 15.8









#35
Methacrylonitrile
Concen: 17.91 ug/l
RT: 4.928 min Scan# 630
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:	41	Resp:	18680
Ion	Ratio	Lower	Upper
41	100		
39	56.0	29.3	87.8
67	69.4	38.3	114.9
52	30.6	15.9	47.6



#36
1,2-Dichloroethane
Concen: 18.21 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:	62	Resp:	32638
Ion	Ratio	Lower	Upper
62	100		
98	9.8	4.6	7.0#





#37

Trichloroethene

Concen: 17.64 ug/l

RT: 7.135 min Scan# 992

Delta R.T. 0.000 min





Lab File: VX021938.D

Acq: 12 May 2021 19:43

Tgt Ion:130 Resp: 23279

Ion Ratio Lower Upper

130	100		
95	102.5	81.1	121.7



#38

Methylcyclohexane

Concen: 17.80 ug/l

RT: 7.391 min Scan# 1034

Delta R.T. 0.006 min




Lab File: VX021938.D


Acq: 12 May 2021 19:43

Tgt Ion: 83 Resp: 34784

Ion Ratio Lower Upper

83	100		
55	74.3	36.4	109.2
98	50.2	25.3	75.9





#39

1,2-Dichloropropane

Concen: 17.68 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.006 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43

Tgt Ion: 63 Resp: 23173

Ion Ratio Lower Upper

63 100





65 27.2 25.2 37.8

Instrument :

MSVOA_X

ClientSampleId :

ICVVX051321



#40

Dibromomethane

Concen: 18.17 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX021938.D

Acq: 12 May 2021 19:43




Tgt Ion: 93 Resp: 16574


Ion Ratio Lower Upper

93 100

95 78.5 0.0 166.0

174 79.7 0.0 170.4









#41
Bromodichloromethane
Concen: 18.10 ug/l
RT: 7.830 min Scan# 1106
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 83 Resp: 32974
Ion Ratio Lower Upper
83 100
85 68.1 52.6 78.8
127 8.3 6.1 9.1



#42
Vinyl Acetate
Concen: 90.30 ug/l
RT: 3.727 min Scan# 433
Delta R.T. -0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 43 Resp: 328483
Ion Ratio Lower Upper
43 100
86 10.5 8.2 12.4









#43
Ethyl Acetate
Concen: 18.72 ug/l
RT: 4.721 min Scan# 596
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 43 Resp: 37059
Ion Ratio Lower Upper
43 100
45 14.1 11.2 16.8



#44
Isopropyl Acetate
Concen: 18.11 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 43 Resp: 59127
Ion Ratio Lower Upper
43 100
61 21.4 17.0 25.4









#45
1,4-Dioxane
Concen: 349.28 ug/l
RT: 7.677 min Scan# 1081
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:	88	Resp:	10181
Ion	Ratio	Lower	Upper
88	100		
43	36.9	0.0	0.0#
58	71.1	0.0	0.0#



#46
Methyl methacrylate
Concen: 18.44 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:	41	Resp:	27914
Ion	Ratio	Lower	Upper
41	100		
69	81.5	42.7	128.1
39	55.7	32.1	96.3









#47
n-amy1 Acetate
Concen: 17.83 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 43 Resp: 48414
Ion Ratio Lower Upper
43 100
55 23.1 18.9 28.3



#48
t-1,3-Dichloropropene
Concen: 18.03 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 75 Resp: 36859
Ion Ratio Lower Upper
75 100
77 36.7 26.8 40.2









#49

cis-1,3-Dichloropropene
Concen: 18.29 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321




Tgt Ion:	75	Resp:	40139
Ion	Ratio	Lower	Upper
75	100		
77	30.6	24.9	37.3
39	46.1	38.6	57.8




#50

1,1,2-Trichloroethane
Concen: 18.04 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:	97	Resp:	23420
Ion	Ratio	Lower	Upper
97	100		
83	85.7	68.2	102.4
85	51.4	44.4	66.6
99	62.2	44.2	66.4









#51
Ethyl methacrylate
Concen: 17.88 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:	69	Resp:	37440
Ion	Ratio	Lower	Upper
69	100		
41	65.7	31.1	93.5
39	37.6	18.3	54.9



#52
1,3-Dichloropropane
Concen: 17.94 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:	76	Resp:	40602
Ion	Ratio	Lower	Upper
76	100		
78	33.4	0.0	64.2









#53
Dibromochloromethane
Concen: 18.54 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:129 Resp: 24186
Ion Ratio Lower Upper
129 100
127 75.6 39.0 116.9



#54
1,2-Dibromoethane
Concen: 18.52 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.007 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:107 Resp: 24696
Ion Ratio Lower Upper
107 100
109 92.4 74.4 111.6









#55
2-Chloroethyl vinyl ether
Concen: 94.00 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 63 Resp: 103898
Ion Ratio Lower Upper
63 100
65 32.0 25.6 38.4
106 28.4 23.2 34.8



#56
Bromoform
Concen: 18.01 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 173 Resp: 14903
Ion Ratio Lower Upper
173 100
175 47.0 39.4 59.2
254 9.0 8.0 12.0









#57
Chlorobenzene-d5
Concen: 30.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:	117	Resp:	97925
Ion	Ratio	Lower	Upper
117	100		
82	58.0	44.9	67.3
119	32.0	25.0	37.6



#58
4-Methyl-2-Pentanone
Concen: 90.19 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.001 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:	43	Resp:	178103
Ion	Ratio	Lower	Upper
43	100		
58	39.2	31.8	47.6
85	18.3	14.9	22.3









#59
2-Hexanone
Concen: 90.72 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 43 Resp: 132682
Ion Ratio Lower Upper
43 100
58 54.4 45.3 67.9
57 18.8 15.0 22.6



#60
4-Bromofluorobenzene
Concen: 29.60 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43


Tgt Ion: 95 Resp: 48110
Ion Ratio Lower Upper
95 100
174 64.5 52.2 78.2
176 61.1 49.3 73.9








#61
Tetrachloroethene
Concen: 17.73 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.005 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321







Tgt Ion:164 Resp: 19690
Ion Ratio Lower Upper
164 100
166 128.8 90.2 135.4
129 107.9 73.4 110.0
131 106.0 75.4 113.0



#62
Toluene
Concen: 17.81 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.007 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 91 Resp: 99848
Ion Ratio Lower Upper
91 100
92 58.5 47.4 71.2









#63
Toluene-d8
Concen: 29.76 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.001 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 98 Resp: 133164
Ion Ratio Lower Upper
98 100
100 66.8 52.6 78.8



#64
Chlorobenzene
Concen: 18.44 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 112 Resp: 61809
Ion Ratio Lower Upper
112 100
114 31.0 25.3 37.9









#65
1,1,1,2-Tetrachloroethane
Concen: 17.74 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:131 Resp: 21708
Ion Ratio Lower Upper
131 100
133 96.1 0.0 187.6
119 65.9 0.0 137.6



#66
Ethyl Benzene
Concen: 18.14 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 91 Resp: 111914
Ion Ratio Lower Upper
91 100
106 30.6 24.6 37.0









#67
m/p-Xylenes
Concen: 35.63 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:106 Resp: 83292
Ion Ratio Lower Upper
106 100
91 203.8 166.0 249.0



#68
o-Xylene
Concen: 18.31 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:106 Resp: 41688
Ion Ratio Lower Upper
106 100
91 213.4 106.0 318.0









#69
Styrene
Concen: 17.73 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:104 Resp: 67543
Ion Ratio Lower Upper
104 100
78 52.8 41.8 62.8
103 53.6 44.1 66.1



#70
Isopropylbenzene
Concen: 17.69 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:105 Resp: 105799
Ion Ratio Lower Upper
105 100
120 26.2 18.9 35.1









#71
1,1,2,2-Tetrachloroethane
Concen: 17.64 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.006 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 83 Resp: 34784
Ion Ratio Lower Upper
83 100
131 8.7 4.7 14.1
85 63.3 30.8 92.3



#72
1,2,3-Trichloropropane
Concen: 23.62 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 75 Resp: 41284
Ion Ratio Lower Upper
75 100
77 32.4 0.0 33.0









#73
Bromobenzene
Concen: 17.29 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:156 Resp: 21907
Ion Ratio Lower Upper
156 100
77 191.9 0.0 363.4
158 99.5 0.0 193.4



#74
n-propylbenzene
Concen: 17.70 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.006 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 91 Resp: 123364
Ion Ratio Lower Upper
91 100
120 22.1 0.0 44.2









#75
2-Chlorotoluene
Concen: 17.94 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 91 Resp: 74083
Ion Ratio Lower Upper
91 100
126 32.2 0.0 63.8



#76
1,3,5-Trimethylbenzene
Concen: 17.90 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 105 Resp: 89390
Ion Ratio Lower Upper
105 100
120 48.7 0.0 96.2









#77
t-1,4-Dichloro-2-butene
Concen: 18.07 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 75 Resp: 12800
Ion Ratio Lower Upper
75 100
53 82.6 40.2 120.6
89 37.7 23.4 70.0



#78
4-Chlorotoluene
Concen: 17.61 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 91 Resp: 83674
Ion Ratio Lower Upper
91 100
126 28.8 0.0 55.8









#79
tert-butylbenzene
Concen: 18.04 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:119 Resp: 86143
Ion Ratio Lower Upper
119 100
91 60.3 0.0 123.0
134 22.5 0.0 45.4



#80
1,2,4-Trimethylbenzene
Concen: 17.78 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:105 Resp: 88713
Ion Ratio Lower Upper
105 100
120 46.1 0.0 91.4









#81
 sec-Butylbenzene
 Concen: 17.57 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX021938.D
 Acq: 12 May 2021 19:43




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX051321


Tgt Ion:105 Resp: 105444
 Ion Ratio Lower Upper
 105 100
 134 19.4 0.0 38.8



#82
 p-Isopropyltoluene
 Concen: 17.51 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX021938.D
 Acq: 12 May 2021 19:43

Tgt Ion:119 Resp: 88105
 Ion Ratio Lower Upper
 119 100
 134 26.6 0.0 50.4
 91 26.2 0.0 51.0









#83
1,3-Dichlorobenzene
Concen: 17.31 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:146 Resp: 41566
Ion Ratio Lower Upper
146 100
111 44.3 22.1 66.5
148 64.6 31.1 93.5



#84
1,4-Dichlorobenzene
Concen: 17.62 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:146 Resp: 42443
Ion Ratio Lower Upper
146 100
111 43.7 21.0 63.0
148 62.7 31.6 95.0









#85
n-Butylbenzene
Concen: 17.06 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion: 91 Resp: 78695
Ion Ratio Lower Upper
91 100
92 54.5 0.0 111.0
134 24.5 0.0 49.8



#86
Hexachloroethane
Concen: 18.13 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 117 Resp: 15161
Ion Ratio Lower Upper
117 100
201 63.2 30.9 92.8









#87
1,2-Dichlorobenzene
Concen: 17.68 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:146 Resp: 41132
Ion Ratio Lower Upper
146 100
111 45.8 22.9 68.8
148 63.9 31.9 95.9



#88
1,2-Dibromo-3-Chloropropane
Concen: 17.15 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion: 75 Resp: 7529
Ion Ratio Lower Upper
75 100
155 74.3 64.6 97.0
157 101.5 75.8 113.6









#89
1,2,4-Trichlorobenzene
Concen: 16.92 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43




Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321


Tgt Ion:180 Resp: 24109
Ion Ratio Lower Upper
180 100
182 93.1 77.8 116.6
145 34.8 30.0 45.0



#90
Hexachlorobutadiene
Concen: 17.94 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:225 Resp: 9892
Ion Ratio Lower Upper
225 100
223 59.9 0.0 128.0
227 59.7 0.0 129.6









#91
Naphthalene
Concen: 17.68 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Instrument :
MSVOA_X
ClientSampleId :
ICVVX051321

Tgt Ion:128 Resp: 98807
Ion Ratio Lower Upper
128 100
127 12.6 10.4 15.6
129 10.8 9.2 13.8



#92
1,2,3-Trichlorobenzene
Concen: 17.68 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX021938.D
Acq: 12 May 2021 19:43

Tgt Ion:180 Resp: 25010
Ion Ratio Lower Upper
180 100
182 95.9 0.0 192.6
145 33.7 0.0 75.8

