

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021939.D
 Acq On : 12 May 2021 20:52
 Operator : JC/MD
 Sample : VX0513WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0513WBL01

Quant Time: May 13 06:25:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 13 06:11:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	20613	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	121362	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	105505	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	47792	30.10	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.33%
60) 4-Bromofluorobenzene	11.085	95	51310	29.30	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	97.67%
63) Toluene-d8	8.653	98	146057	30.29	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.97%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
Data File : VX021939.D
Acq On : 12 May 2021 20:52
Operator : JC/MD
Sample : VX0513WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0513WBL01

Quant Time: May 13 06:25:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 13 06:11:35 2021
Response via : Initial Calibration

