

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021948.D
 Acq On : 13 May 2021 00:19
 Operator : JC/MD
 Sample : M2359-02 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-2

Manual Integrations
 APPROVED

MMDadoda
 5/13/2021 4:07:39 PM

Quant Time: May 13 06:28:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 13 06:11:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.916	128	17773	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	104182	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.061	117	92329	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	42013	30.69	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.30%
60) 4-Bromofluorobenzene	11.085	95	45245	29.53	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	98.43%
63) Toluene-d8	8.653	98	126414	29.96	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.87%
Target Compounds						
25) Chloroform	5.099	83	2416m	1.13	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
Data File : VX021948.D
Acq On : 13 May 2021 00:19
Operator : JC/MD
Sample : M2359-02 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14B-2

Manual Integrations
APPROVED

MMDadoda
5/13/2021 4:07:39 PM

Quant Time: May 13 06:28:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 13 06:11:35 2021
Response via : Initial Calibration

