

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021964.D
 Acq On : 13 May 2021 14:09
 Operator : JC/MD
 Sample : M2383-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 14 04:36:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79387	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	176291	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182790	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83639	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	71539	54.76	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.52%	
35) Dibromofluoromethane	5.397	113	60249	49.87	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.74%	
50) Toluene-d8	8.653	98	240124	52.08	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.16%	
62) 4-Bromofluorobenzene	11.085	95	103333	56.79	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
Data File : VX021964.D
Acq On : 13 May 2021 14:09
Operator : JC/MD
Sample : M2383-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: May 14 04:36:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

