

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021965.D
 Acq On : 13 May 2021 14:30
 Operator : JC/MD
 Sample : M2383-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 14 04:36:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	79812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	173900	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	180505	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81516	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	69018	52.55	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.10%
35) Dibromofluoromethane	5.397	113	59785	50.17	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.34%
50) Toluene-d8	8.653	98	233252	51.29	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.58%
62) 4-Bromofluorobenzene	11.085	95	101720	56.67	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.34%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
Data File : VX021965.D
Acq On : 13 May 2021 14:30
Operator : JC/MD
Sample : M2383-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: May 14 04:36:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

