

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021966.D
 Acq On : 13 May 2021 14:52
 Operator : JC/MD
 Sample : M2383-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-01-05112021

Quant Time: May 14 04:36:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	82799	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	179716	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83029	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	70407	51.67	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.34%
35) Dibromofluoromethane	5.403	113	60551	49.16	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.32%
50) Toluene-d8	8.653	98	242634	51.62	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.24%
62) 4-Bromofluorobenzene	11.085	95	103292	55.69	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.38%
Target Compounds						
16) Acetone	2.380	43	4306	7.05	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
Data File : VX021966.D
Acq On : 13 May 2021 14:52
Operator : JC/MD
Sample : M2383-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-01-05112021

Quant Time: May 14 04:36:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

