

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021967.D
 Acq On : 13 May 2021 15:13
 Operator : JC/MD
 Sample : M2383-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-02-05122021

Quant Time: May 14 04:36:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	80978	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	176638	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	185800	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84836	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	69818	52.39	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.78%
35) Dibromofluoromethane	5.397	113	60437	49.93	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.86%
50) Toluene-d8	8.653	98	238580	51.65	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.30%
62) 4-Bromofluorobenzene	11.085	95	104315	57.22	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.44%
Target Compounds						
16) Acetone	2.380	43	4043	6.76	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
Data File : VX021967.D
Acq On : 13 May 2021 15:13
Operator : JC/MD
Sample : M2383-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-02-05122021

Quant Time: May 14 04:36:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

