

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021976.D
 Acq On : 13 May 2021 18:36
 Operator : JC/MD
 Sample : PB136362ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136362ZHE#02

Quant Time: May 14 04:40:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	79820	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	175011	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	184480	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82384	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	64268	48.92	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.84%
35) Dibromofluoromethane	5.391	113	56514	47.12	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.24%
50) Toluene-d8	8.653	98	220683	48.22	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.44%
62) 4-Bromofluorobenzene	11.085	95	96061	53.18	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.36%
Target Compounds						
16) Acetone	2.380	43	9689	16.45	ug/l	# 88
17) Carbon Disulfide	2.514	76	11870	3.94	ug/l	96
20) Methylene Chloride	2.794	84	11993	9.99	ug/l	93
43) Isopropyl Acetate	6.348	43	14891	3.93	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
Data File : VX021976.D
Acq On : 13 May 2021 18:36
Operator : JC/MD
Sample : PB136362ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136362ZHE#02

Quant Time: May 14 04:40:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

