

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021977.D
 Acq On : 13 May 2021 18:59
 Operator : JC/MD
 Sample : PB136362ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136362ZHE#03

Quant Time: May 14 04:40:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	77770	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	170634	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	175430	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78526	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	62153	48.56	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.12%		
35) Dibromofluoromethane	5.397	113	55102	47.12	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.24%		
50) Toluene-d8	8.653	98	213539	47.85	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.70%		
62) 4-Bromofluorobenzene	11.085	95	92049	52.27	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.54%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	9224	16.07	ug/l	97
17) Carbon Disulfide	2.514	76	8626	2.94	ug/l	99
20) Methylene Chloride	2.794	84	12545	10.71	ug/l	95
43) Isopropyl Acetate	6.348	43	15377	4.16	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
Data File : VX021977.D
Acq On : 13 May 2021 18:59
Operator : JC/MD
Sample : PB136362ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136362ZHE#03

Quant Time: May 14 04:40:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

