


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051321\
 Data File : VX021931.D
 Acq On : 12 May 2021 16:56
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu May 13 06:11:35 2021



AutoFind: Averaged scan 1635 to 1649; Bkg corrected with scan 1634

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	1453	PASS
75	95	30	60	49.7	3968	PASS
95	95	100	100	100.0	7987	PASS
96	95	5	9	6.8	546	PASS
173	174	0.00	2	0.6	30	PASS
174	95	50	100	63.0	5030	PASS
175	174	5	9	8.1	405	PASS
176	174	95	101	97.8	4919	PASS
177	176	5	9	5.9	291	PASS