

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028669.D
 Acq On : 13 May 2022 11:46
 Operator : JC/MD
 Sample : VX0513WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0513WBL01

Quant Time: May 16 02:18:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	228876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	427768	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	410271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184984	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	168196	53.611	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.220%
35) Dibromofluoromethane	5.391	113	143358	51.150	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.300%
50) Toluene-d8	8.653	98	518385	49.059	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.120%
62) 4-Bromofluorobenzene	11.085	95	192906	48.878	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028669.D
Acq On : 13 May 2022 11:46
Operator : JC/MD
Sample : VX0513WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0513WBL01

Quant Time: May 16 02:18:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

