

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028676.D
 Acq On : 13 May 2022 14:33
 Operator : JC/MD
 Sample : N2794-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB01-20220510

Quant Time: May 16 02:21:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	222294	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.769	114	408740	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	387790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171443	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162792	53.425	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.840%
35) Dibromofluoromethane	5.391	113	136347	50.914	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.820%
50) Toluene-d8	8.653	98	494604	48.987	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.980%
62) 4-Bromofluorobenzene	11.079	95	180232	47.792	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028676.D
Acq On : 13 May 2022 14:33
Operator : JC/MD
Sample : N2794-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB01-20220510

Quant Time: May 16 02:21:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

