

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028685.D
 Acq On : 13 May 2022 18:03
 Operator : JC/MD
 Sample : N2770-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M-21519

Quant Time: May 16 02:23:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	215830	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	408962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	391710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162260	54.845	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 109.680%		
35) Dibromofluoromethane	5.391	113	137312	51.246	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 102.500%		
50) Toluene-d8	8.653	98	497241	49.222	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 98.440%		
62) 4-Bromofluorobenzene	11.085	95	181485	48.099	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 96.200%		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	30572	24.078	ug/l	93
51) 4-Methyl-2-Pentanone	8.580	43	16677	3.850	ug/l	90
68) m/p-Xylenes	10.305	106	2248	0.388	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028685.D
Acq On : 13 May 2022 18:03
Operator : JC/MD
Sample : N2770-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M-21519

Quant Time: May 16 02:23:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

