

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028688.D
 Acq On : 13 May 2022 19:13
 Operator : JC/MD
 Sample : N2861-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 16 02:24:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	214355	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	405009	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	394061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161168	54.851	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.700%	
35) Dibromofluoromethane	5.391	113	135072	50.902	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.800%	
50) Toluene-d8	8.653	98	497626	49.741	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.480%	
62) 4-Bromofluorobenzene	11.079	95	180206	48.226	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028688.D
Acq On : 13 May 2022 19:13
Operator : JC/MD
Sample : N2861-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 16 02:24:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

