

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028690.D
 Acq On : 13 May 2022 20:00
 Operator : JC/MD
 Sample : N2830-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1594

Quant Time: May 16 02:25:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	224025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	421492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	402478	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	180264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166026	54.065	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 108.120%			
35) Dibromofluoromethane	5.391	113	141189	51.127	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.260%			
50) Toluene-d8	8.653	98	508326	48.823	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.640%			
62) 4-Bromofluorobenzene	11.085	95	185751	47.766	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.540%			
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	12675	11.848	ug/l	99
16) Acetone	2.374	43	293114	222.403	ug/l	91
17) Carbon Disulfide	2.514	76	9194	1.530	ug/l	99
25) 2-Butanone	4.556	43	131025	63.336	ug/l	94
51) 4-Methyl-2-Pentanone	8.580	43	6306	1.413	ug/l	91
52) Toluene	8.720	92	47078	6.139	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028690.D
Acq On : 13 May 2022 20:00
Operator : JC/MD
Sample : N2830-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1594

Quant Time: May 16 02:25:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

