

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028700.D
 Acq On : 14 May 2022 00:39
 Operator : JC/MD
 Sample : PB144781ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144781ZHE#01

Quant Time: May 16 02:30:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	232012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	430025	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	423267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184281	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	168310	52.922	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.840%	
35) Dibromofluoromethane	5.391	113	141744	50.309	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.620%	
50) Toluene-d8	8.653	98	536007	50.460	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.920%	
62) 4-Bromofluorobenzene	11.079	95	192931	48.628	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.260%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	38549	28.242	ug/l	92
20) Methylene Chloride	2.788	84	12348	4.312	ug/l #	91
43) Isopropyl Acetate	6.348	43	21923	3.170	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028700.D
Acq On : 14 May 2022 00:39
Operator : JC/MD
Sample : PB144781ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144781ZHE#01

Quant Time: May 16 02:30:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

