

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028702.D
 Acq On : 14 May 2022 01:25
 Operator : JC/MD
 Sample : PB144781ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144781ZHE#03

Quant Time: May 16 02:30:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	225057	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	428927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	425381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	187006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166285	53.901	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.800%	
35) Dibromofluoromethane	5.391	113	141736	50.435	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.880%	
50) Toluene-d8	8.652	98	528874	49.916	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.840%	
62) 4-Bromofluorobenzene	11.079	95	194058	49.037	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.080%	
Target Compounds						Qvalue
16) Acetone	2.379	43	39349	29.720	ug/l	90
20) Methylene Chloride	2.794	84	13458	4.845	ug/l	95
43) Isopropyl Acetate	6.348	43	21642	3.138	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028702.D
Acq On : 14 May 2022 01:25
Operator : JC/MD
Sample : PB144781ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144781ZHE#03

Quant Time: May 16 02:30:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

