

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028705.D
 Acq On : 14 May 2022 02:35
 Operator : JC/MD
 Sample : PB144781ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144781ZHE#06

Quant Time: May 16 02:31:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	225292	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	432515	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	433288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	219005	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	169176	54.781	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.560%	
35) Dibromofluoromethane	5.385	113	140962	49.744	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.480%	
50) Toluene-d8	8.653	98	536050	50.174	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.340%	
62) 4-Bromofluorobenzene	11.079	95	201419	50.475	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.940%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	39631	29.901	ug/l	95
20) Methylene Chloride	2.788	84	13239	4.761	ug/l	91
43) Isopropyl Acetate	6.348	43	21867	3.144	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028705.D
Acq On : 14 May 2022 02:35
Operator : JC/MD
Sample : PB144781ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144781ZHE#06

Quant Time: May 16 02:31:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

