

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028709.D
 Acq On : 14 May 2022 04:08
 Operator : JC/MD
 Sample : N2824-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COM-2

Quant Time: May 16 02:33:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	244147	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	451815	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	428979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	219944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	176627	52.777	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.560%	
35) Dibromofluoromethane	5.385	113	77524	26.189	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 52.380%#	
50) Toluene-d8	8.653	98	552313	49.488	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.980%	
62) 4-Bromofluorobenzene	11.079	95	199155	47.775	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.560%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	63515	44.221	ug/l	94
18) Methyl Acetate	2.703	43	2969	0.825	ug/l	99
20) Methylene Chloride	2.788	84	14974	4.969	ug/l	90
25) 2-Butanone	4.562	43	10574	4.690	ug/l	93
51) 4-Methyl-2-Pentanone	8.580	43	5765	1.205	ug/l	91
95) Naphthalene	13.774	128	25810	1.651	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028709.D
Acq On : 14 May 2022 04:08
Operator : JC/MD
Sample : N2824-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COM-2

Quant Time: May 16 02:33:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

