

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028712.D
 Acq On : 14 May 2022 05:17
 Operator : JC/MD
 Sample : PB144804TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144804TB

Quant Time: May 16 02:34:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	241313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	450675	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	500556	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190172	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	178116	53.847	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.700%	
35) Dibromofluoromethane	5.391	113	147704	50.022	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.040%	
50) Toluene-d8	8.653	98	552744	49.652	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.300%	
62) 4-Bromofluorobenzene	11.079	95	220208	52.960	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 105.920%	
Target Compounds						
						Qvalue
16) Acetone	2.373	43	40425	28.475	ug/l	95
20) Methylene Chloride	2.788	84	10984	3.688	ug/l #	82
43) Isopropyl Acetate	6.348	43	22905	3.161	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028712.D
Acq On : 14 May 2022 05:17
Operator : JC/MD
Sample : PB144804TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144804TB

Quant Time: May 16 02:34:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

