

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
 Data File : VX028716.D
 Acq On : 14 May 2022 06:50
 Operator : JC/MD
 Sample : N2814-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HEATER-WATER

Quant Time: May 16 02:35:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	495965	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	466669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	219182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	194147	52.443	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.880%
35) Dibromofluoromethane	5.391	113	161956	49.840	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.680%
50) Toluene-d8	8.653	98	595762	48.629	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.260%
62) 4-Bromofluorobenzene	11.079	95	223018	48.737	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.480%
Target Compounds						
11) Tert butyl alcohol	2.946	59	10138	5.087	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028716.D
Acq On : 14 May 2022 06:50
Operator : JC/MD
Sample : N2814-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HEATER-WATER

Quant Time: May 16 02:35:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

