



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051322\
Data File : VX028694.D
Acq On : 13 May 2022 21:56
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Title : SW846 8260
Last Update : Thu May 12 14:58:53 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	9751	PASS
75	95	30	60	53.3	26949	PASS
95	95	100	100	100.0	50581	PASS
96	95	5	9	6.9	3469	PASS
173	174	0.00	2	1.3	492	PASS
174	95	50	100	77.6	39243	PASS
175	174	5	9	7.6	2972	PASS
176	174	95	101	96.0	37667	PASS
177	176	5	9	6.3	2374	PASS