

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
 Data File : VX041478.D
 Acq On : 13 May 2024 09:38
 Operator : JC/MD
 Sample : VX0513MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0513MBL01

Quant Time: May 14 01:20:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	189794	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	280455	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137185	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104459	50.557	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	101.120%	
35) Dibromofluoromethane	5.379	113	86918	45.733	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	91.460%	
50) Toluene-d8	8.647	98	328198	54.559	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	109.120%	
62) 4-Bromofluorobenzene	11.079	95	121861	52.332	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	104.660%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
Data File : VX041478.D
Acq On : 13 May 2024 09:38
Operator : JC/MD
Sample : VX0513MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0513MBL01

Quant Time: May 14 01:20:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

