

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
 Data File : VX041486.D
 Acq On : 13 May 2024 12:46
 Operator : JC/MD
 Sample : P2455-01DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-03B-90-051024DL

Quant Time: May 14 01:30:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	162685	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	243281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119156	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91062	51.417	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	102.840%	
35) Dibromofluoromethane	5.385	113	74270	45.049	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	90.100%	
50) Toluene-d8	8.647	98	280333	53.723	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	107.440%	
62) 4-Bromofluorobenzene	11.079	95	105625	52.291	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	104.580%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	32241	25.228	ug/l	Qvalue 88
12) 1,1-Dichloroethene	2.313	96	1078	0.855	ug/l #	74
27) cis-1,2-Dichloroethene	4.489	96	3464	2.077	ug/l	81
44) Trichloroethene	7.123	130	66352	41.347	ug/l	90
64) Tetrachloroethene	9.275	164	751	0.448	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
Data File : VX041486.D
Acq On : 13 May 2024 12:46
Operator : JC/MD
Sample : P2455-01DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-051024DL

Quant Time: May 14 01:30:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

