

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
 Data File : VX041488.D
 Acq On : 13 May 2024 13:32
 Operator : JC/MD
 Sample : P2445-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-04B-92.5-050924

Quant Time: May 14 01:31:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	172317	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	252536	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	244221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122605	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96498	51.441	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.880%
35) Dibromofluoromethane	5.385	113	78897	46.102	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.200%
50) Toluene-d8	8.647	98	295852	54.619	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.240%
62) 4-Bromofluorobenzene	11.079	95	111832	53.335	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.660%
Target Compounds						
30) Chloroform	5.105	83	1959	0.701	ug/l	# 78
44) Trichloroethene	7.129	130	623	0.374	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
Data File : VX041488.D
Acq On : 13 May 2024 13:32
Operator : JC/MD
Sample : P2445-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-04B-92.5-050924

Quant Time: May 14 01:31:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

