

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
 Data File : VX041491.D
 Acq On : 13 May 2024 14:41
 Operator : JC/MD
 Sample : P2449-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1011UMR-RB240508-01

Quant Time: May 14 01:32:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148391	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	218804	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83872	51.919	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.840%			
35) Dibromofluoromethane	5.385	113	68249	46.028	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.060%			
50) Toluene-d8	8.647	98	253202	53.952	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 107.900%			
62) 4-Bromofluorobenzene	11.079	95	92564	50.951	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.900%			
Target Compounds						
16) Acetone	2.392	43	9019	8.906	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
Data File : VX041491.D
Acq On : 13 May 2024 14:41
Operator : JC/MD
Sample : P2449-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1011UMR-RB240508-01

Quant Time: May 14 01:32:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

