

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
 Data File : VX041492.D
 Acq On : 13 May 2024 15:04
 Operator : JC/MD
 Sample : P2449-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1011UMR-TB240508

Quant Time: May 14 01:33:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	153256	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	228372	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219561	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	112910	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87649	52.535	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.060%
35) Dibromofluoromethane	5.385	113	70619	45.631	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.260%
50) Toluene-d8	8.647	98	263451	53.784	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.560%
62) 4-Bromofluorobenzene	11.079	95	100872	53.198	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.400%
Target Compounds						
16) Acetone	2.392	43	7964	7.038	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
Data File : VX041492.D
Acq On : 13 May 2024 15:04
Operator : JC/MD
Sample : P2449-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1011UMR-TB240508

Quant Time: May 14 01:33:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

