

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
 Data File : VX041498.D
 Acq On : 13 May 2024 17:22
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 14 01:35:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156195	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	230996	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198984	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107891	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87740	51.600	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.200%
35) Dibromofluoromethane	5.385	113	70632	45.121	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.240%
50) Toluene-d8	8.647	98	224002	45.211	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.420%#
62) 4-Bromofluorobenzene	11.079	95	98468	51.340	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
Data File : VX041498.D
Acq On : 13 May 2024 17:22
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 14 01:35:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

