

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
 Data File : VX041499.D
 Acq On : 13 May 2024 17:45
 Operator : JC/MD
 Sample : P2449-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1011UMR-TB240508

Quant Time: May 14 01:36:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	153873	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	229443	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87577	52.281	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.560%	
35) Dibromofluoromethane	5.385	113	70234	45.170	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	90.340%	
50) Toluene-d8	8.647	98	221138	44.935	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	89.860%#	
62) 4-Bromofluorobenzene	11.079	95	79964	41.974	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	83.940%	
Target Compounds						
16) Acetone	2.386	43	9159	8.640	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051324\
Data File : VX041499.D
Acq On : 13 May 2024 17:45
Operator : JC/MD
Sample : P2449-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1011UMR-TB240508

Quant Time: May 14 01:36:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

