

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051325\
 Data File : VX046154.D
 Acq On : 13 May 2025 10:29
 Operator : JC/MD
 Sample : VX0513MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0513MBL01

Quant Time: May 14 01:34:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125679	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48833	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	63167	53.384	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.760%
35) Dibromofluoromethane	5.373	113	46670	51.568	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.140%
50) Toluene-d8	8.647	98	156874	50.081	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.160%
62) 4-Bromofluorobenzene	11.079	95	59880	49.836	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051325\
Data File : VX046154.D
Acq On : 13 May 2025 10:29
Operator : JC/MD
Sample : VX0513MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0513MBL01

Quant Time: May 14 01:34:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

