

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051325\
 Data File : VX046156.D
 Acq On : 13 May 2025 11:16
 Operator : JC/MD
 Sample : Q1993-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-20250508

Quant Time: May 14 01:35:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	65629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132457	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54742	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	67850	55.454	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.900%
35) Dibromofluoromethane	5.379	113	49334	51.722	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.440%
50) Toluene-d8	8.647	98	168956	51.178	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.360%
62) 4-Bromofluorobenzene	11.079	95	65184	51.474	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051325\
Data File : VX046156.D
Acq On : 13 May 2025 11:16
Operator : JC/MD
Sample : Q1993-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-20250508

Quant Time: May 14 01:35:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

