

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051325\
 Data File : VX046174.D
 Acq On : 13 May 2025 18:18
 Operator : JC/MD
 Sample : Q2017-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-J

Quant Time: May 14 01:45:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	65625	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130004	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63786	52.136	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.280%		
35) Dibromofluoromethane	5.385	113	46683	49.866	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.740%		
50) Toluene-d8	8.647	98	164013	50.618	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.240%		
62) 4-Bromofluorobenzene	11.079	95	65992	53.095	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 106.200%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	26256	53.524	ug/l	97
18) Methyl Acetate	2.709	43	2543	2.234	ug/l	98
20) Methylene Chloride	2.782	84	2759	2.935	ug/l	95
43) Isopropyl Acetate	6.348	43	13442	5.670	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051325\
Data File : VX046174.D
Acq On : 13 May 2025 18:18
Operator : JC/MD
Sample : Q2017-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-J

Quant Time: May 14 01:45:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

