

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051325\
 Data File : VX046175.D
 Acq On : 13 May 2025 18:42
 Operator : JC/MD
 Sample : Q2007-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-636-COMP-10

Quant Time: May 14 01:45:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	137266	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67427	53.196	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.400%		
35) Dibromofluoromethane	5.379	113	49474	50.052	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.100%		
50) Toluene-d8	8.647	98	175314	51.244	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.480%		
62) 4-Bromofluorobenzene	11.079	95	67774	51.644	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 103.280%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20755	40.839	ug/l	97
18) Methyl Acetate	2.709	43	2397	2.033	ug/l	97
20) Methylene Chloride	2.782	84	3927	4.032	ug/l	87
43) Isopropyl Acetate	6.348	43	10281	4.107	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051325\
Data File : VX046175.D
Acq On : 13 May 2025 18:42
Operator : JC/MD
Sample : Q2007-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-636-COMP-10

Quant Time: May 14 01:45:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

