

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016225.D
 Acq On : 14 May 2020 13:00
 Operator : JC/SP
 Sample : L2576-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FRAC-TANK-F05194

Quant Time: May 15 09:16:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	254413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	420584	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	417817	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	215530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	158097	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
35) Dibromofluoromethane	5.47	113	124851	46.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.72%	
50) Toluene-d8	8.70	98	520164	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	11.12	95	210358	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
Target Compounds						
16) Acetone	2.42	43	8297	5.554	ug/l	96
17) Carbon Disulfide	2.56	76	10815	1.160	ug/l	96
20) Methylene Chloride	2.84	84	70632	22.937	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016225.D
Acq On : 14 May 2020 13:00
Operator : JC/SP
Sample : L2576-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-F05194

Quant Time: May 15 09:16:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

