

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016230.D
 Acq On : 14 May 2020 14:53
 Operator : JC/SP
 Sample : L2643-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW126-200513

Quant Time: May 15 02:17:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	256870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	429398	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	423550	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	220258	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	160360	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
35) Dibromofluoromethane	5.47	113	128931	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
50) Toluene-d8	8.70	98	530941	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
62) 4-Bromofluorobenzene	11.12	95	214917	53.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.82%	
Target Compounds						
4) Vinyl Chloride	1.39	62	7607	2.258	ug/l	99
16) Acetone	2.42	43	2442	1.619	ug/l	91
17) Carbon Disulfide	2.56	76	3617	0.245	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	55528	17.070	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016230.D
Acq On : 14 May 2020 14:53
Operator : JC/SP
Sample : L2643-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW126-200513

Quant Time: May 15 02:17:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

