

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016231.D
 Acq On : 14 May 2020 15:16
 Operator : JC/SP
 Sample : L2643-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW098-200513

Quant Time: May 15 02:19:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	254618	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	424162	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	418475	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	212531	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	160344	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
35) Dibromofluoromethane	5.47	113	128015	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	8.70	98	528235	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
62) 4-Bromofluorobenzene	11.12	95	207878	52.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.56%	
Target Compounds						
16) Acetone	2.42	43	3339	2.233	ug/l	95
27) cis-1,2-Dichloroethene	4.57	96	8084	2.507	ug/l	85
30) Chloroform	5.18	83	1287	0.252	ug/l	88
44) Trichloroethene	7.19	130	38120	8.783	ug/l	95
64) Tetrachloroethene	9.32	164	1871	0.364	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016231.D
Acq On : 14 May 2020 15:16
Operator : JC/SP
Sample : L2643-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW098-200513

Quant Time: May 15 02:19:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

