

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016233.D
 Acq On : 14 May 2020 16:01
 Operator : JC/SP
 Sample : L2643-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW601-200513

Quant Time: May 15 02:21:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	248246	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	415251	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	406459	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	155862	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
35) Dibromofluoromethane	5.47	113	124235	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
50) Toluene-d8	8.70	98	520867	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
62) 4-Bromofluorobenzene	11.12	95	203296	52.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.46%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.32	50	2286	0.749	ug/l	100
16) Acetone	2.42	43	5890	4.041	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	17813	5.666	ug/l	84
44) Trichloroethene	7.20	130	24419	5.747	ug/l	91
64) Tetrachloroethene	9.32	164	1352	0.271	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016233.D
Acq On : 14 May 2020 16:01
Operator : JC/SP
Sample : L2643-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW601-200513

Quant Time: May 15 02:21:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

