

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016234.D
 Acq On : 14 May 2020 16:24
 Operator : JC/SP
 Sample : L2643-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS004MW010-200513

Quant Time: May 15 09:17:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	241461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	402515	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	402831	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	205619	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	151249	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
35) Dibromofluoromethane	5.47	113	121704	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
50) Toluene-d8	8.70	98	502363	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
62) 4-Bromofluorobenzene	11.12	95	201369	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
Target Compounds						
4) Vinyl Chloride	1.39	62	43065	13.599	ug/l	99
16) Acetone	2.42	43	2268	1.600	ug/l #	78
27) cis-1,2-Dichloroethene	4.57	96	20814	6.807	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016234.D
Acq On : 14 May 2020 16:24
Operator : JC/SP
Sample : L2643-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS004MW010-200513

Quant Time: May 15 09:17:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

