

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016236.D
 Acq On : 14 May 2020 17:09
 Operator : JC/SP
 Sample : L2643-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS004MW153-200513

Quant Time: May 15 02:26:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	237685	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	396039	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	396454	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	197555	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	147025	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
35) Dibromofluoromethane	5.47	113	116563	46.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.92%	
50) Toluene-d8	8.70	98	492630	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	11.12	95	196061	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
Target Compounds						
16) Acetone	2.42	43	5127	3.674	ug/l	95
27) cis-1,2-Dichloroethene	4.56	96	1586	0.527	ug/l	47
30) Chloroform	5.18	83	1509	0.317	ug/l #	77
44) Trichloroethene	7.19	130	14246	3.515	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016236.D
Acq On : 14 May 2020 17:09
Operator : JC/SP
Sample : L2643-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS004MW153-200513

Quant Time: May 15 02:26:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

