

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016237.D
 Acq On : 14 May 2020 17:31
 Operator : JC/SP
 Sample : L2643-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW750-200513

Quant Time: May 15 04:04:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	249717	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	418185	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	417651	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	214056	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	156210	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
35) Dibromofluoromethane	5.48	113	126026	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
50) Toluene-d8	8.70	98	521457	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	11.12	95	207182	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
Target Compounds						
16) Acetone	2.42	43	2124	1.449	ug/l	91
44) Trichloroethene	7.19	130	3656	0.854	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016237.D
Acq On : 14 May 2020 17:31
Operator : JC/SP
Sample : L2643-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW750-200513

Quant Time: May 15 04:04:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

